



Instructions for Use  
**ShurFit® ACIF 2C Anterior Cervical Interbody Fusion System**

Caution: Federal (USA) law restricts this device to sale by or on the order of a physician

**DISCLAIMER OF WARRANTY AND LIMITATION OF REMEDY**

There is no express or implied warranty, including any implied warranty of merchantability or fitness for a particular purpose, on **Precision Spine®** product(s) described in this publication. Under no circumstances shall **Precision Spine** be liable for any direct, incidental or consequential damages other than as expressly provided by specific law. No person has the authority to bind **Precision Spine** to any representation or warranty except as specifically set forth herein.

Descriptions or specifications in **Precision Spine** printed matter, including this publication, are meant solely to generally describe the product at the time of manufacture and do not constitute any express warranties.

**DEVICE DESCRIPTION**

The **ShurFit ACIF 2C** Anterior Cervical Interbody Fusion System consists of implants with various heights to accommodate individual patient anatomy and graft material size. It is implanted from the anterior approach. It is to be packed with autogenous bone graft to facilitate fusion. The device is intended to provide mechanical support to the implanted level until biologic fusion is achieved. All components are manufactured from medical grade polyetheretherketone (PEEK, per ASTM F2026) and Tantalum (ASTM-F560), and coated with Commercially Pure Titanium in compliance with ASTM F1580 and Hydroxyapatite in compliance with ASTM F1185-03. The products are supplied clean and "STERILE".

**INDICATIONS:**

The **ShurFit ACIF 2C** Anterior Cervical Interbody Fusion System is indicated for use in skeletally mature patients with degenerative disc disease (DDD) of the cervical spine at one disc level. DDD is defined as discogenic pain with degeneration of the disc confirmed by history and radiographic studies. **ShurFit ACIF 2C** Anterior Cervical Interbody Fusion System implants are used to facilitate fusion in the cervical spine and are placed via an anterior approach at the C-3 to C-7 disc levels using autograft bone. Patients should have at least six weeks of non-operative treatment prior to treatment with an intervertebral fusion device. The device should be used with supplemental fixation.

**PRECAUTIONS:**

Based on the dynamic testing results, the physician should consider the levels of implantation, patient weight, patient activity level, other patient conditions, etc., which may impact on the performance of the intervertebral body fusion device.

The implantation of the intervertebral body fusion device should be performed only by experienced spinal surgeons with specific training in the use of this device because this is a technically demanding procedure presenting a risk of serious injury to the patient.

All system implants are single-use only. Reuse of the device may result in the following:

1. Infection
2. Fracture / mechanical failure of the device
3. Inability to properly engage surgical instrumentation
4. Pyrogenic reaction

**CONTRAINDICATIONS:**

The **ShurFit ACIF 2C** Anterior Cervical Interbody Fusion System contraindications include, but not limited to:

1. Prior fusion at the level(s) to be treated
2. Any condition not described in the indications for use
3. Previous vascular approach
4. Iliofemoral arteriosclerosis
5. Morbid obesity
6. Mental illness
7. Pregnancy
8. Local infection or inflammation

9. Any case needing to mix metals from different components
10. Any patient unwilling to cooperate with postoperative instructions
11. All cases not stated in the indications
12. Reuse, or multiple uses

#### **POTENTIAL ADVERSE AFFECTS:**

The following potential adverse effects associated with the procedure have been shown to occur with the use of similar spinal systems. All patients considered candidates for fusion should be informed concerning the pathogenesis of their spinal abnormality, the rationale for fusion with instrumentation, and the potential adverse affects. The following are potential adverse effects, but not limited to:

1. Loss of proper spinal curvature, correction, height, and/or reduction
2. Infection
3. Non-Union or delayed union
4. Foreign body reaction to the implants
5. Hemorrhaging
6. Loss of neurological function, dural tear, pain, and/or discomfort
7. Bone graft fracture, vertebral body fracture or discontinued growth of fused at, above and/or below the surgery level
8. Bending, loosening, fracture, disassembly, slippage and/or migration of the components
9. Pain or discomfort
10. Change in mental status
11. Bursitis
12. Bone loss and/or bone fracture due to stress shielding
13. Inability to resume activities of normal daily activities
14. Revision surgery
15. Death

#### **WARNINGS:**

The following are warnings of this device.

1. Patients with previous spinal surgery at the level(s) to be treated may have different clinical outcomes compared to those without previous surgery.
2. Potential risks identified with the use of this device system, which may require additional surgery, include device component fracture, loss of fixation, non-union, fracture of the vertebrae, necrosis of the bone, neurological injury, and/or vascular or visceral injury.
3. The benefit of spinal fusions utilizing any interbody fusion device has not been adequately established in patients with stable spines.
4. Patient selection and compliance will greatly affect the results. Patients suffering from obesity, malnutrition, and/or poor bone quality are poor candidates for spinal fusion. Patients who smoke, or abuse alcohol, are poor candidates for spinal fusion.
5. Patients who smoke should be advised of the consequences of the fact that an increased incidence of non-union has been reported with patients who smoke.
6. The implants and instruments are provided non-sterile and must be cleaned and sterilized before use. Device components should be sterilized using one of the noted validated sterilization cycle parameters.
7. A successful result is not always achieved in every surgical case due to many extenuating circumstances. This is especially true in spinal surgeries where other patient conditions may compromise the results.
8. Never reuse an internal fixation device under any circumstances.
9. This device is not intended to be the sole means of spinal support. The **ShurFit® ACIF 2C** Anterior Cervical Interbody Fusion System must be used with additional anterior and/or posterior instrumentation to augment stability.
10. Only surgeons trained and experienced in spinal decompression and bone grafting techniques should use the **ShurFit ACIF 2C** Anterior Cervical Interbody Fusion System. Preoperative and operating procedures, including knowledge of surgical techniques and proper selection and placement of the implants are essential considerations in the utilization of this device.
11. Physician note: Although the physician is the learned intermediary, the important medical information given in this document should be conveyed to the patient.
12. Do not reuse implants. Discard used, damaged, or otherwise suspect implants. **AN IMPLANT SHOULD NEVER BE RE-USED.** Any implant, once used, should be discarded. Even though it appears undamaged, it may have small defects and internal stress patterns that may lead to failure. These Single Use devices have not been designed to undergo or withstand any form of alteration, such as disassembly, cleaning or re-sterilization, after a single patient use. Reuse can potentially compromise device performance and patient safety.

**PREOPERATIVE:**

1. The surgeon should only consider utilizing the **ShurFit® ACIF 2C** Anterior Cervical Interbody Fusion System with those patients who meet the criteria in **Indications**.
2. The surgeon should avoid utilizing this device with those patients who have **Contraindications**.
3. The surgeon should make sure that all implants and instruments are unpacked, sterilized, and available prior to surgery.
4. The implants and instruments are provided non-sterile and must be cleaned and sterilized before use.
5. Implants and instruments should be inspected for surface flaws and scratches and should not be used in the presence thereof.
6. The surgeon should have a complete understanding of the surgical technique, design rationale, indications and contraindications.
7. The surgeon should have a complete understanding of the surgical technique manual.

**INTRAOPERATIVE:**

1. The instructions in any available applicable surgical technique manual should be carefully followed.
2. Damage to the nerves will cause loss of neurological functions. Extreme caution should be taken to avoid the spinal cord and nerve roots at all times.
3. Careful use of the implants and instruments should be taken. Misuse of the components may cause injury to the patient or operative personnel.
4. Bone graft should be packed inside the device prior to insertion and around the device after insertion. Bone graft must be placed in the area to be fused. The bone graft must extend from the upper to the lower vertebrae to be fused.
5. Notching and scratching of implants should be avoided.
6. The **ShurFit ACIF 2C** Anterior Cervical Interbody Fusion System should be supported by anterior and/or posterior stabilization devices.

**POSTOPERATIVE:**

1. The physician’s postoperative directions, warnings to the patient and the corresponding patient’s compliance are extremely important.
2. For best possible results, patients should be counseled to avoid lifting, twisting, physical activities, smoking, consuming alcohol, and any other activity that would compromise or delay the healing process.
3. The patient should be warned about the limitation of bending at the point of spinal fusion.
4. The surgeon should instruct the patient regarding the amount and time frame after surgery of any weight bearing activity. The increased risk of bending, dislocation, and/or breakage of the implant devices, as well as an undesired surgical result are consequences of any type of early or excessive weight bearing, vibration motion, fall, jolts, or other movements preventing proper healing and/or fusion.
5. The removal of implants should be properly disposed of and are not to be reused under any circumstance.

**STERILIZATION:**

The **ShurFit ACIF 2C** Anterior Cervical Interbody Fusion System is supplied as sterile, therefore cleaning and sterilization is not required. Do not use if found damaged or expired, and contact the manufacturer. All other components, including instruments, are supplied clean and non-sterile and must be sterilized prior to use. Remove all packaging before sterilization. Instruments should be autoclave sterilized using one of the following validated cycle parameters.

Method	Cycle Type	Sterilization Temperature	Minimum Exposure Time
Steam	Gravity Displacement	270°F (132°C)	15 minutes
Steam	Pre-vacuum	270°F (132°C)	4 minutes

To assure maintenance of sterility we recommend

- Utilization of a minimum drying time of 20 minutes in accordance with ANSI/AAMI ST79:2010, *Comprehensive guide to steam sterilization and sterility assurance in health facilities*.
- For USA: Only use FDA cleared sterilization wraps to enclose the sterilization tray.

**MAGNETIC RESONANCE ENVIRONMENT:**

The **ShurFit 2C** Anterior Cervical Interbody Fusion System has not been evaluated for safety and compatibility in the Magnetic Resonance environment. The **ShurFit ACIF 2C** Anterior Cervical Interbody Fusion System has not been tested for heating or migration in the Magnetic Resonance environment.

**STORAGE INSTRUCTIONS:**

All products should be stored in a cool dry place.

#### HOW SUPPLIED:

The **ShurFit® ACIF 2C** Anterior Cervical Interbody Fusion System is supplied sterile. All other components and specialized instruments are supplied non-sterile in a container suitable for steam sterilization or individually packaged as replacement product. All components and instruments may be purchased independently.

#### CARE AND HANDLING:

- All torque handles should be returned to the manufacturer for recalibration every six months.
- Please refer to ASTM standards such as F1744-96, "Standard Guide for Care and Handling of Stainless Steel Surgical Instruments" for additional information.
- Surgical instruments are subject to wear with normal usage. Instruments which have experienced extensive use or excessive force are susceptible to fracture. Surgical instruments should only be used for their intended purpose.
- Precision Spine recommends that all instruments be visually inspected for wear and disfigurement, as well as tested, to ensure instruments are functioning properly prior to use. If instruments are discolored, have loose screws/pins, are out of alignment, are cracked or have other irregularities, **DO NOT USE**.

#### CLEANING AND DECONTAMINATION:

- These instructions are to be followed prior to initial use and reprocessing of all instruments.
- Reprocessing the instruments, using the methods described herein, will not limit the useful life of the instruments. The useful life of the product is typically determined by wear and damage due to use.
- Transport trays should be considered reusable devices, inspected for visible soils and must be cleaned.
- **WARNING:** The following Cleaning and Sterilization instructions have been validated. Failure to follow all steps may result in an improperly cleaned and sterilized instrument (Non-Sterile).
- **CAUTION:** In order to preserve optimal efficiency and safety of the instruments, the following instructions must be followed.
  - The use of metallic brushes, scrub pads or other articles that are likely to damage the instrument must be avoided.
  - Chemicals, such as chlorine or soda, as well as organic or ammoniated acids or solvents (ex. Acetone) that are likely to damage the instrument, must not be used.
  - Mercurial solutions are not recommended, as they corrode metal parts.
  - If applicable, disassemble instruments prior to Cleaning and Sterilization. Articulated instruments must be opened in order to allow the cleaning of all interstices.
  - Immediately after the surgical procedure, disallowing organic debris to dry on the instruments, remove as much debris as possible from each instrument using a water moistened gauze pad or wipe, exchanging the gauze pad or wipe as it becomes soiled.
  - Prepare a neutral pH enzymatic cleaning solution per the manufacturer's instructions with warm tap water (35-40° C).
  - Immerse the instruments in the cleaning solution for a minimum of 10 minutes, activating any mechanisms 5X, so the enzymatic cleaner contacts all mated surfaces. Thoroughly scrub all instruments with a soft bristle cleaning brush while immersed in the enzymatic cleaning solutions. Be sure that thorough scrubbing also includes any lumens with an appropriately sized brush that contacts all surfaces. Change the soak solution after each utilization or if grossly soiled.
  - Rinse the instruments in warm tap water (35-40° C) for at least one minute.
  - Transfer the instruments into fresh enzymatic cleaning solution. Sonicate the instruments while immersed in the cleaning solution for a minimum of 15 minutes.
  - Thoroughly rinse all instruments and lumens with warm running water (35-40° C) for at least one minute each until flushing water runs clear. Use a hose or water jet to rinse any lumens, holes, or complex interfaces. Perform a second rinse with DI water, again using a hose or water jet to rinse any lumens, holes or complex interfaces.
  - Dry with a sterile gauze, clean cloth and/or clean compressed air. Inspect instruments for cleanliness, function and residual moisture. Any device that is not visually clean must be reprocessed.

#### LUBRICATION:

To protect instruments from staining and rusting during sterilization and storage, they should be lubricated with a water-soluble, preserved lubricant after each cleaning. Since effective ultrasonic cleaning removes all lubricant, relubrication is important. The lubricant should contain a chemical preservative to prevent bacterial growth in the lubricant bath. The bath solution should be made with demineralized water. A lubricant containing a rust inhibitor helps prevent electrolytic corrosion of points and edges. Immediately after cleaning, instrument should be immersed for 30 seconds and allowed to drain off, not wiped off. A lubricant film will remain through the sterilization to protect them during storage.

**SPECIAL NOTE FOR TORQUE LIMITING HANDLES:**

(This note applies only to customers who purchase Torque Limiting Handles.)

The following are suggested guidelines for calibration cycles of Torque Limiting Handles. Note that these are general recommendations only and users are encouraged to determine specific calibration cycles for each product depending on their particular situation or usage. Return product after six months of use, or after 150 autoclave cycles, or after approximately 3,000 actuations (Clicks), whichever comes first.

**MATERIAL SPECIFICATION:**

The **ShurFit® ACIF 2C** Anterior Cervical Interbody Fusion System is manufactured from medical grade Polyetheretherketone (PEEK, per ASTM F2026) and Tantalum as per ASTM-F560, and coated with Commercially Pure Titanium in compliance with ASTM F1580, and Hydroxyapatite in compliance with ASTM F1185-03. All other components are manufactured from medical grade stainless steel, titanium, titanium alloy, or medical grade plastics described by such standards as ASTM F-138, ASTM F-136, ASTM D-6394 or ISO 5832-3.

**CLINICAL HISTORY:**

These instructions for use are based upon current experience. The physician may wish to vary the procedure in accordance with clinical judgment.

**PRODUCT COMPLAINTS:**

Any complaint or dissatisfaction with product quality, performance, labeling, and/or safety should be reported to **Precision Spine®**. If any of the implants or instruments "malfunction" (i.e. do not meet any of their performance specifications or does not perform as intended), and/or are suspected to have caused or contributed to the death or serious injury of the patient, **Precision Spine** should be notified immediately by phone, fax or written correspondence. When filing a complaint, please provide the product description, product number, lot number, your name and address, and the nature of the complaint.

**ADDITIONAL INFORMATION:**

The surgical technique guide for the implantation of the **ShurFit ACIF 2C** Anterior Cervical Interbody Fusion System is available upon request. If further information is required, please contact the manufacturer.



**Precision Spine, Inc.**

2050 Executive Drive  
Pearl, MS 39208  
USA  
Phone: 1-601-420-4244  
Toll Free: 1-888-241-4773  
Fax: 1-601-420-5501



**Emergo Europe**

Prinsessegracht 20  
2514 AP The Hague  
The Netherlands  
Tel: (31) (0) 70 345-8570

<b>ShurFit® ACIF 2C Anterior Cervical Interbody Fusion System</b>				
				<b>RX only</b>
SEE PACKAGE INSERT FOR LABELING LIMITATIONS	STERILIZED BY IRRADIATION	YR/MONTH/DAY EXPIRATION DATE	SINGLE USE ONLY	SALE BY PHYSICIAN PRESCRIPTION USA ONLY
MANUFACTURED BY				